

**CERTIFICATE OF ANALYSIS**

# Canine Parvovirus (CPV) IgM

FA Positive Control

Catalog No.:	PC-IFA-CPV-M
Lot:	P151029-002
Expiration:	24 November 2019

Quality Control Method: IFA using CPV 12-well slide (catalog no. SLD-IFA-CPV) and anti-canine IgM FITC conjugate (catalog no. CJ-F-CANM-AP-1ML or 10ML).

Specific Reaction: 1-4+ signal and no background with the control at neat, endpoint titer reached at 1/2.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 24 November 2015

Pattern Of Fluorescence: Single cells with cytoplasmic and nuclear fluorescence.

Technical Data Sheet Version: Version # 1

**FOR IN VITRO LABORATORY USE ONLY.**

H:\Quality\COAs and TDS\COA (VMRD)\Positive Controls\CPV\CPV IgM\PC-IFA-CPV-M\_P151029-002\_191124.docx

30 November 2015