

CERTIFICATE OF ANALYSIS

Canine Parvovirus (CPV) IgM

FA Positive Control

Catalog No.:	PC-IFA-CPV-M
Lot:	P200916-001
Expiration:	23 September 2024

Quality Control Method: IFA using CPV 12-well slide (catalog no. SLD-IFA-CPV) and anti-canine IgM FITC conjugate (catalog no. CJ-F-CANM-AP-10ML).

Specific Reaction: 2–4+ signal with the control at neat, trace–1 background. Endpoint signal was reached at 1/8 dilution.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Single cells with cytoplasmic and nuclear fluorescence.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

24 September 2020