

CERTIFICATE OF ANALYSIS

Canine Parvovirus (CPV) IgM

FA Positive Control

| | |
|--------------|------------------|
| Catalog No.: | PC-IFA-CPV-M |
| Lot: | P201010-001 |
| Expiration: | 13 November 2024 |

Quality Control Method: IFA using CPV 12-well slide (catalog no. SLD-IFA-CPV) and anti-canine IgM FITC conjugate (catalog no. CJ-F-CANM-AP-10ML).

Specific Reaction: 2-4+ signal with the control at neat, no background. Endpoint signal was reached at 1/4 dilution.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Single cells with cytoplasmic and nuclear fluorescence.

Technical Data Sheet Version: 1