

CERTIFICATE OF ANALYSIS

*Ehrlichia canis*

FA Positive Control

Catalog No.:	PC-IFA-EC
Lot:	P190812-006
Expiration:	3 October 2023

Quality Control Method: Indirect FA using *Ehrlichia canis* 12-well slide (catalog no. SLD-IFA-EC), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 1-3+ signal with the control at neat, trace background. Endpoint is 1+ at 1/4 dilution.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

Technical Data Sheet Version: 1

FOR IN VITRO LABORATORY USE ONLY.

14 October 2019