

CERTIFICATE OF ANALYSIS

# Feline Infectious Peritonitis Virus Type 1 (FIP-1)


FA Positive Control

Catalog No.:	PC-IFA-FIP1
Lot:	P231211
Expiration:	08 January 2032

**Quality Control Method:** Indirect FA using FIP-1 12-well slides (catalog no. SLD-IFA-FIP1), negative control (catalog no. NC-IFA-FIP1), and anti-feline IgG affinity purified FITC conjugate (catalog no. CJ-F-FELG-AP-10ML).

**Specific Reaction:** 2–4+ signal with the control at neat. No background. Endpoint is greater than four 2-fold dilutions.

**Other Comments:** NA

**QC Manager:** 

David Hyatt

**Pattern Of Fluorescence:** Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

19 January 2024