

CERTIFICATE OF ANALYSIS

## Feline Immunodeficiency Virus (FIV)

FA Positive Control

Catalog No.:	PC-IFA-FIV
Lot:	P220401-002
Expiration:	25 May 2026

**Quality Control Method:** IFA using FIV 12-well slide (catalog no. SLD-IFA-FIV), negative control (catalog no. NC-IFA-FIV), and anti-feline IgG affinity purified FITC conjugate (catalog no. CJ-F-FELG-AP-10ML).

**Specific Reaction:** 2–3+ signal with the control at neat. No background. Endpoint signal was reached at 1/16 dilution.

**Other Comments:** NA

**QC Manager:** 

David Hyatt

**Pattern Of Fluorescence:** Syncytia and individual cells with granular and diffuse cytoplasmic fluorescence.

Technical Data Sheet Version: 1

FOR IN VITRO LABORATORY USE ONLY.

06 June 2022