

CERTIFICATE OF ANALYSIS

Feline Herpesvirus Type 1 (FHV-1/FVR)

FA Positive Control

Catalog No.:	PC-IFA-FVR
Lot:	P170803-001
Expiration:	17 August 2021

Quality Control Method: Indirect FA using FHV-1/FVR 12-well slide (catalog no. SLD-IFA-FVR), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

Specific Reaction: 2-4+ signal with the control at neat, no background. Endpoint is at greater than a 1/16 dilution.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 24 August 2017

Pattern of Fluorescence: Individual rounded cells with cytoplasmic and nuclear fluorescence. Some small syncytia.

Technical Data Sheet Version: Version #1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Positive Controls\FHV-FVR\PC-IFA-FVR_P170803-001_210817.docx

30 August 2017