

CERTIFICATE OF ANALYSIS

Feline Herpesvirus Type 1 (FHV-1/FVR)


FA Positive Control

Catalog No.:	PC-IFA-FVR
Lot:	P201211-001
Expiration:	13 January 2025

Quality Control Method: Indirect FA using FHV-1/FVR 12-well slide (catalog no. SLD-IFA-FVR), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-10ML).

Specific Reaction: 3–4+ signal with the control at neat, no background. Endpoint is at a dilution greater than 1/16.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Individual rounded cells with cytoplasmic and nuclear fluorescence. Some small syncytia.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

15 January 2021