

CERTIFICATE OF ANALYSIS

Borrelia burgdorferi (Lyme Disease)

FA Positive Control

Catalog No.:	PC-IFA-LD
Lot:	P200403-001
Expiration:	18 March 2032

Quality Control Method: Indirect FA using *Borrelia burgdorferi* 12-well slide (catalog no. SLD-IFA-LD), negative control (catalog no. NC-IFA-LD), and anti-Canine IgG affinity purified FITC conjugate (catalog no. CJ-F-CANG-AP-10ML).

Specific Reaction: 3–4+ signal with the control at neat. No background. Endpoint signal was reached at 1/8 dilution.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Spirochetes with uniform surface fluorescence.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

20 March 2024