

CERTIFICATE OF ANALYSIS

Porcine Reproductive and Respiratory Syndrome Virus North American Strain (PRRSV-NA)


FA Positive Control

Catalog No.:	PC-IFA-PRRS-NA
Lot:	P151009-001
Expiration:	28 March 2031

Quality Control Method: Indirect FA using PRRSV in-house 2-well slides, negative control (Catalog no. NC-IFA-PRRS), and anti-porcine IgG affinity purified FITC conjugate (catalog no. CJ-F-PORG-AP-10ML).

Specific Reaction: 2–4+ signal with the positive control at neat. No background. Trace–1+ endpoint reached in 3 two-fold dilutions.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Granular and diffuse fluorescence on cytoplasm, membrane, and occasional nuclear and cytoplasmic inclusions.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

5 April 2023