

CERTIFICATE OF ANALYSIS

*Rickettsia rickettsii* (RMSF)

FA Positive Control

Catalog No.:	PC-IFA-RMSF
Lot:	P150309-003
Expiration:	19 March 2019

Quality Control Method: Indirect FA using *R. rickettsii* 12-well slide (catalog no. SLD-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 3-4+ fluorescence at neat with an endpoint titer of 1+ at 1/4.

Other Comments: NA

QC Manager: *Torrey Beeson* Torrey Beeson

QC Release Date: 20 March 2015

Pattern Of Fluorescence: Fluorescent coccobacilli-like structures among ovine red blood cells.  
Some clumped.

Technical Data Sheet Version: Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Quality VMRD\COAs and TDS\COA (VMRD)\Positive Controls\RMSF\PC-IFA-RMSF\_P150309-003\_190319.docx

23 March 2015