

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Positive Control

Catalog No.:	PC-IFA-RMSF
Lot:	P170209-002
Expiration:	20 February 2021

Quality Control Method: Indirect FA using *R. rickettsii* 12-well slide (catalog no. SLD-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 2-3+ signal with the positive control at neat, no background. Endpoint of Tr-1+ reached at 1/4 dilution.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 20 February 2017

Pattern of Fluorescence: Fluorescent coccobacilli-like structures among ovine red blood cells. Some clumped.

Technical Data Sheet Version: Version #1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\Positive Controls\RMSF\PC-IFA-RMSF_P170209-002_210220.docx

22 February 2017