

CERTIFICATE OF ANALYSIS

*Rickettsia rickettsii* (RMSF)

FA Positive Control

Catalog No.:	PC-IFA-RMSF
Lot:	P170511-001
Expiration:	14 December 2024

**Quality Control Method:** Indirect FA using *R. rickettsii* 12-well slide (catalog no. SLD-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-10ML).

**Specific Reaction:** 3–4+ signal with the positive control at neat, no background. Endpoint of trace–1+ was reached at a 1/8 dilution.

**Other Comments:** NA

**QC Manager:**



David Hyatt

**Pattern of Fluorescence:** Fluorescent coccobacilli-like structures among ovine red blood cells. Some clumped.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Positive Controls\RMSF\PC-IFA-RMSF\_P170511-001\_241214.docx

15 December 2020