

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)


FA Positive Control

Catalog No.:	PC-IFA-RMSF
Lot:	P211110-001
Expiration:	11 November 2025

Quality Control Method: Indirect FA using *R. rickettsii* 12-well slide (catalog no. SLD-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-AP-10ML).

Specific Reaction: 3+ signal with the control at neat, no background. Endpoint signal was reached at 1/8 dilution.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Fluorescent coccobacilli-like structures among ovine red blood cells. Some clumped.

Technical Data Sheet Version: 1