

CERTIFICATE OF ANALYSIS

Equine Infectious Anemia Virus (EIAV)

ELISA Strong Positive Reference Serum

Catalog No.:	RS-EIA-ES-0.5ML
Lot:	P230519-001
Expiration:	14 June 2025

Quality Control Method: Sample was tested in the VMRD EIAV ELISA (catalog no. 290-1) and AGID (catalog no. 400-200) according to the kit inserts.

ELISA Specific Reaction: Average OD is 0.970

Sample	OD Values		Average OD
QC serum	0.962	0.978	0.970
+ Control	0.214 0.215	0.216 0.216	0.215

AGID Specific Reaction: Positive

Other Comments: NA

QC Manager: 

David Hyatt

Technical Data Sheet Version: 1

FOR IN VITRO LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Reference Samples\RS-EIA-EW-0.5ML_P230519-001_250614.docx

27 June 2023