Veterinary Medical
Research & Development

CERTIFICATE OF ANALYSIS

Bovine Coronavirus (BCV)

FA Control Slide

Catalog No.:	SLD-FAC-BCV
Lot:	P171020-010
Expiration:	03 November 2021

Quality Control Method: Direct FA using BCV FITC conjugate (catalog no. CJ-F-BCV-1ML or 10ML).

Specific Reaction: 1-3+ signal on the positive well and negative on the negative well with no background. There are 0 to 50 infected cells per high-power field.

Other Comments: This lot of slides exhibits inconsistent distribution of infected cells.

QC Manager: Will Thurman

QC Release Date: 07 November 2017

Pattern Of Fluorescence: Granular cytoplasmic and nuclear fluorescence mainly on single cells but also some small syncytia.

Technical Data Sheet Version: Version 2

FOR *IN VITRO* LABORATORY USE ONLY.

PO Box 502
Pullman, WA 99163 USA

P: 509.334.5815
F: 509.332.5356

techserve@vmrd.com
www.vmrd.com