

**CERTIFICATE OF ANALYSIS**

# Bovine Parvovirus (BPV)

FA Control Slide

Catalog No.:	SLD-FAC-BPV
Lot:	P221025-001
Expiration:	04 November 2030

Quality Control Method: Direct FA using BPV FITC conjugate (catalog no. CJ-F-BPV-10ML).

Specific Reaction: 3–4+ signal on the positive well and negative on the negative well. Trace background. There are 0–10 infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Ground glass and/or granular cytoplasmic fluorescence with occasional bright and diffuse nuclear fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

10 November 2022