

CERTIFICATE OF ANALYSIS

Canine Parvovirus (CPV)

FA Control Slide

Catalog No.:	SLD-FAC-CPV
Lot:	P200617-001
Expiration:	08 April 2032

Quality Control Method: Direct FA using CPV direct conjugate (catalog no. CJ-F-CPV-MAB-10ML).

Specific Reaction: 3–4+ signal on the positive well and negative on the negative well. No background. There are more than 50 infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Single cells with cytoplasmic and nuclear fluorescence.

Technical Data Sheet Version: 3

FOR *IN VITRO* LABORATORY USE ONLY.

18 April 2024