

CERTIFICATE OF ANALYSIS

Equine Herpesvirus Type 1 (EHV-1/ERV)

FA Control Slide

Catalog No.:	SLD-FAC-ERV
Lot:	P190812-004
Expiration:	04 March 2032

Quality Control Method: Direct FA using EHV-1/ERV conjugate (catalog no. CJ-F-ERV-10ML).

Specific Reaction: 2–4+ signal on the positive well and negative on the negative well. No background. There are 0–25 infected cells and or syncytia per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Strong undifferentiated cytoplasmic fluorescence. Single cells with occasional nuclear inclusions and some small syncytia. Some intense fluorescence of dying rounded cells.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

11 March 2024