

CERTIFICATE OF ANALYSIS

Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Control Slide

Catalog No.:	SLD- FAC-FIP1
Lot:	P190830-005
Expiration:	08 August 2031

Quality Control Method: Direct FA using FIP conjugate (catalog no. CJ-F-FIP-10ML).

Specific Reaction: 2–3+ signal on the positive well and negative on the negative well. No background. There are 0 to too numerous to count infected cells and/or syncytia per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

29 August 2023