

CERTIFICATE OF ANALYSIS

FELINE HERPESVIRUS TYPE 1 (FHV-1)

FA Control Slide

Catalog No.:	SLD-FAC-FVR
Lot:	P191125-001
Expiration:	17 March 2031

Quality Control Method: Direct FA using FHV-1/FVR FA conjugate (catalog no. CJ-F-FVR-1ML or 10ML).

Specific Reaction: 1–3+ signal on the positive well and negative on the negative well. Trace background. There are 0–20 plaques and/or infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Individual rounded cells with cytoplasmic and nuclear fluorescence. Some small syncytia

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

31 March 2023