

CERTIFICATE OF ANALYSIS

Rabies Recombinant Nucleoprotein (rNP)

FA Control Slide

Catalog No.:	SLD-FAC-RAB
Lot:	P220515-003
Expiration:	25 May 2026

Quality Control Method: Direct FA using in-house Rabies FITC conjugate.

Specific Reaction: 2-4+ signal on the positive well and negative on the negative well. No background. There are 0-25 infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Smooth and particulate cytoplasmic fluorescence of single and multinucleated cells.

Technical Data Sheet Version: 3

FOR *IN VITRO* LABORATORY USE ONLY.

06 June 2022