

CERTIFICATE OF ANALYSIS

Reovirus (REO)

FA Control Slide

Catalog No.:	SLD-FAC-REO
Lot:	P200210-002
Expiration:	18 June 2024

Quality Control Method: Direct FA using REO conjugate (catalog no. CJ-F-REO-10ML).

Specific Reaction: 2-4+ signal on the positive well and negative on the negative well with no background. There are 0-25 infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Individual cells or plaques with diffuse cytoplasmic fluorescence and characteristic bright cytoplasmic inclusions ranging in size from medium to large.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

21 July 2020