

CERTIFICATE OF ANALYSIS

*Anaplasma phagocytophilum*

FA Substrate Slide

Catalog No.:	SLD-IFA-AP
Lot:	P190102-004
Expiration:	16 April 2023

**Quality Control Method:** Indirect FA using *Anaplasma phagocytophilum* positive control (catalog no. PC-IFA-AP), negative control (catalog no. NC-IFA-AP), anti-equine IgG FITC conjugate (catalog no. CJ-F-EQUG-1ML or 10ML).

**Specific Reaction:** 2 – 4+ signal with the positive control, no background. Negative with the negative control with trace background. There are 0 - 60 infected cells per high-power field.

**Other Comments:** NA

**QC Manager:** Chris Cay



**QC Release Date:** 16 April 2019

**Pattern Of Fluorescence:** Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

Technical Data Sheet Version: Version # 2

FOR IN VITRO LABORATORY USE ONLY.

15 May 2019