

CERTIFICATE OF ANALYSIS

Anaplasma phagocytophilum

FA Substrate Slide

Catalog No.:	SLD-IFA-AP
Lot:	P200203-001
Expiration:	21 April 2024

Quality Control Method: Indirect FA using *Anaplasma phagocytophilum* positive control (catalog no. PC-IFA-AP), negative control (catalog no. NC-IFA-AP), anti-equine IgG FITC conjugate (catalog no. CJ-F-EQUG-10ML).

Specific Reaction: 3-4+ signal with the positive control and negative with the negative control with trace-1+ background. There are 0-25 infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

Technical Data Sheet Version: 2

FOR IN VITRO LABORATORY USE ONLY.

11 May 2020