

CERTIFICATE OF ANALYSIS

Anaplasma phagocytophilum


FA Substrate Slide

Catalog No.:	SLD-IFA-AP
Lot:	P220331-002
Expiration:	12 July 2030

Quality Control Method: Indirect FA using *Anaplasma phagocytophilum* positive control (catalog no. PC-IFA-AP), negative control (catalog no. NC-IFA-AP), anti-equine IgG FITC conjugate (catalog no. CJ-F-EQUG-10ML).

Specific Reaction: 4+ signal with the positive control and negative with the negative control. No background. There are 5–30 infected cells per high-power field.

Other Comments: NA

QC Manager: 
David Hyatt

Pattern Of Fluorescence: Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

21 July 2022