

CERTIFICATE OF ANALYSIS

Anaplasma phagocytophilum

FA Substrate Slide

Catalog No.:	SLD-IFA-AP
Lot:	P220331-004
Expiration:	19 August 2030

Quality Control Method: Indirect FA using *Anaplasma phagocytophilum* positive control (catalog no. PC-IFA-AP), negative control (catalog no. NC-IFA-AP), anti-equine IgG FITC conjugate (catalog no. CJ-F-EQUG-10ML).

Specific Reaction: 2–4+ signal with the positive control and negative with the negative control. No background. Organisms per high-power field are too numerous to count.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

30 August 2022