

CERTIFICATE OF ANALYSIS

Bluetongue Virus (BTV)

FA Substrate Slide

Catalog No.:	SLD-IFA-BTV
Lot:	P170804-001
Expiration:	30 August 2021

Quality Control Method: Indirect FA using BTV positive control (catalog no. PC-IFA-BTV), negative control (catalog no. NC-IFA-BTV), and anti-bovine IgG_{1,2} affinity purified FITC conjugate (catalog no. CJ-F-BOVG-AP-1ML or 10ML).

Specific Reaction: 2-4+ signal with the positive control, negative with the negative control and no background. 0-25 infected cells and or syncytia per high powered field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 31 August 2017

Pattern of Fluorescence: Multiple buckshot inclusions with dusty cytoplasmic fluorescence in individual cells and occasional syncytia.

Technical Data Sheet Version: Version #2

FOR IN VITRO LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\8-, 10-, 12-well Slides\BTV\SLD-IFA-BTV_P170804-001_210829.docx

06 September 2017