

CERTIFICATE OF ANALYSIS

## Bluetongue Virus (BTV)

FA Substrate Slide

Catalog No.:	SLD-IFA-BTV
Lot:	P220729-002
Expiration:	25 August 2030

**Quality Control Method:** Indirect FA using BTV positive control (catalog no. PC-IFA-BTV), negative control (catalog no. NC-IFA-BTV), and anti-bovine IgG<sub>1,2</sub> affinity purified FITC conjugate (catalog no. CJ-F-BOVG-AP-10ML).

**Specific Reaction:** 2–4+ signal with the positive control and negative with the negative control. No background. Infected cells per high-power field range from 35 to too numerous to count.

**Other Comments:** NA

**QC Manager:**



David Hyatt

**Pattern Of Fluorescence:** Multiple buckshot inclusions with dusty cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

30 August 2022