

CERTIFICATE OF ANALYSIS

Clostridium spp. 4-way

FA Substrate Slide

Catalog No.:	SLD-IFA-C4
Lot:	P180716-002
Expiration:	15 June 2026

Quality Control Method: Direct FA using conjugates: *C. chauvoei* (catalog no. CJ-F-CCO-10ML), *C. novyi* (catalog no. CJ-F-CNO-10ML), *C. septicum* (catalog no. CJ-F-CSE-10ML), and *C. sordellii* (catalog no. CJ-F-CSO-10ML).

Specific Reaction: *C. chauvoei* had 3–4+ signal, *C. novyi* had 2–3+ signal, *C. septicum* had 3–4+ signal, and *C. sordellii* had 2–3+ signal. There are greater than 10 organisms per high powered field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Individual rod-shaped bacteria. Some clumped bacteria. Flagella fluorescence alone is not indicative of a positive signal.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

22 June 2022