

CERTIFICATE OF ANALYSIS

Brucella canis

FA Substrate Slide

Catalog No.:	SLD-IFA-CB
Lot:	P200817-005
Expiration:	06 November 2024

Quality Control Method: IFA using *Brucella canis* positive control (catalog no. PC-IFA-CB), negative control (catalog no. NC-IFA-CB), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-10ML).

Specific Reaction: 3–4+ signal with the positive control at neat, no background. Organisms per high-power field are too numerous to count.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Brightly staining short bacilli against uniform background of ovine red blood cells.

Technical Data Sheet Version: 2

FOR IN VITRO LABORATORY USE ONLY.

13 November 2020