

CERTIFICATE OF ANALYSIS

Canine Distemper Virus (CDV)

FA Substrate Slide

Catalog No.:	SLD-IFA-CDV
Lot:	P160816-002
Expiration:	17 October 2020

Quality Control Method: Indirect FA using CDV IgG positive control (catalog no. PC-IFA-CDV-G), anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML), CDV IgM positive control (catalog no. PC-IFA-CDV-M), and anti-canine IgM affinity purified FITC conjugate (catalog no. CJ-F-CANM-AP-1ML or 10ML).

Specific Reaction: 2-4+ signal with the IgG positive control. 1-3+ with the IgM positive control. Negative with the diluent and no background. There are 0-50 infected cells per high powered field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 18 October 2016

Pattern of Fluorescence: Individual cells with inclusion bodies and some plasma membrane fluorescence.

Technical Data Sheet Version: Version #2

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\8-, 10-, 12-well Slides\CDV\SLD-IFA-CDV_P160816-002_201017.docx

21 October 2016