

CERTIFICATE OF ANALYSIS

Canine Distemper Virus (CDV)

FA Substrate Slide

Catalog No.:	SLD-IFA-CDV
Lot:	P191004-002
Expiration:	21 January 2024

Quality Control Method: Indirect FA using CDV IgG positive control (catalog no. PC-IFA-CDV-G), anti-canine IgG affinity purified FITC conjugate (catalog no. CJ-F-CANG-AP-10ML), CDV IgM positive control (catalog no. PC-IFA-CDV-M), and anti-canine IgM affinity purified FITC conjugate (catalog no. CJ-F-CANM-AP-10ML).

Specific Reaction: 2-4+ signal with the IgG positive control, No background. 1-4+ signal with the IgM positive control, trace background. Negative with the diluent control, no background. There are 0-50 infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Individual cells with inclusion bodies and some plasma membrane fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

17 February 2020