

CERTIFICATE OF ANALYSIS

## Canine Parvovirus (CPV)

FA Substrate Slide

Catalog No.:	SLD-IFA-CPV
Lot:	P220726-001
Expiration:	23 September 2030

**Quality Control Method:** Indirect FA using CPV IgG positive control (catalog no. PC-IFA-CPV-G), CPV in house negative control, CPV IgM positive control (catalog no. PC-IFA-CPV-M), anti-canine IgG affinity purified FITC conjugate (catalog no. CJ-F-CANG-AP-10ML), and anti-canine IgM affinity purified FITC conjugate (catalog no. CJ-F-CANM-AP-10ML).

**Specific Reaction:** 3–4+ signal with the IgG positive control and 4+ signal with the IgM positive control. Negative with the negative control. No background. There are 0–50 infected cells per high-power field.

**Other Comments:** NA

**QC Manager:**



David Hyatt

**Pattern Of Fluorescence:** Single cells with cytoplasmic and nuclear fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

06 October 2022