

CERTIFICATE OF ANALYSIS

Ehrlichia canis

FA Substrate Slide

Catalog No.:	SLD-IFA-EC
Lot:	P150720-001
Expiration:	26 August 2019

Quality Control Method: Indirect FA using *Ehrlichia canis* positive control (catalog no. PC-IFA-EC), negative control (catalog no. NC-IFA-EC), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 3-4+ signal with the positive control, negative with the negative control. 0-50 infected cells per high powered field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 26 August 2015

Pattern Of Fluorescence: Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

Technical Data Sheet Version: Version # 1

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality\COAs and TDS\COA (VMRD)\8-, 10-, 12-well Slides\E. canis\SLD-IFA-EC_P150720-001_190826.docx

31 August 2015