

CERTIFICATE OF ANALYSIS

*Ehrlichia canis*

FA Substrate Slide

Catalog No.:	SLD-IFA-EC
Lot:	P171204-001
Expiration:	22 December 2021

Quality Control Method: Indirect FA using *Ehrlichia canis* positive control (catalog no. PC-IFA-EC), negative control (catalog no. NC-IFA-EC), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 3-4+ signal with positive control. Negative with trace background with negative control. 0-50 infected cells per high powered field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 22 December 2017

Pattern Of Fluorescence: Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

Technical Data Sheet Version: Version # 2

FOR IN VITRO LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\8-, 10-, 12-well Slides\E. canis\SLD-IFA-EC\_P171204-001\_211222.docx

22 December 2017