

CERTIFICATE OF ANALYSIS

*Ehrlichia canis*

FA Substrate Slide

Catalog No.:	SLD-IFA-EC
Lot:	P180803-001
Expiration:	05 October 2022

**Quality Control Method:** Indirect FA using *Ehrlichia canis* positive control (catalog no. PC-IFA-EC), negative control (catalog no. NC-IFA-EC), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

**Specific Reaction:** 3-4 + signal with the positive control with trace background and negative with the negative control with trace background. There are 0 - 50 infected cells per high-power field.

**Other Comments:** NA

**QC Manager:** Chris Cay



**QC Release Date:** 08 October 2018

**Pattern Of Fluorescence:** Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

**Technical Data Sheet Version:** Version # 2

FOR IN VITRO LABORATORY USE ONLY.

25 October 2018