

CERTIFICATE OF ANALYSIS

Ehrlichia canis

FA Substrate Slide

Catalog No.:	SLD- IFA-EC
Lot:	P190201-003
Expiration:	5 November 2023

Quality Control Method: Indirect FA using *Ehrlichia canis* positive control (catalog no. PC-IFA-EC), negative control (catalog no. NC-IFA-EC), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-10ML).

Specific Reaction: 2-4+ signal with trace background with the positive control. Negative with trace-1+ background with the negative control. There are 0-50 infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Brightly fluorescent cytoplasmic inclusion bodies (morulae), clustered elementary bodies (initial bodies), and free elementary bodies.

Technical Data Sheet Version: 2

FOR IN VITRO LABORATORY USE ONLY.

27 November 2019