

**CERTIFICATE OF ANALYSIS**

# Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P150402-001
Expiration:	19 June 2019

**Quality Control Method:** Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

**Specific Reaction:** 2-4+ fluorescence on the positive control and trace on the negative control with no background. There are 0 to 25 infected cells and/or syncytia per high-power field.

**Other Comments:** NA

**QC Manager:** *Torrey Beeson* Torrey Beeson

**QC Release Date:** 19 June 2015

**Pattern Of Fluorescence:** Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Quality\COAs and TDS\COA (VMRD)\8-, 10-, 12-well Slides\FIP-1\SLD-IFA-FIP1\_P150402-001\_190619.docx

22 June 2015