

CERTIFICATE OF ANALYSIS

Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P151221-002
Expiration:	11 February 2020

Quality Control Method: Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

Specific Reaction: 3-4+ signal with the positive control and trace signal with the negative control with no background. There are 0-50 infected cells and/or syncytia per high-power field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 12 February 2016

Pattern Of Fluorescence: Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: Version #1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\8-, 10-, 12-well Slides\FIP-1\SLD-IFA-FIP1_P151221-002_200211.docx

24 February 2016