

CERTIFICATE OF ANALYSIS

# Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P160531-001
Expiration:	29 July 2020

**Quality Control Method:** Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

**Specific Reaction:** 3-4+ signal with the positive control, trace signal with the negative control, no background. 0-25 infected cells and/or syncytia per high powered field.

**Other Comments:** NA

**QC Manager:**  Will Thurman

**QC Release Date:** 29 July 2016

**Pattern of Fluorescence:** Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: Version # 2

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\8-, 10-, 12-well Slides\FIP-1\SLD-IFA-FIP1\_P160531-001\_200729.docx

03 August 2016