

CERTIFICATE OF ANALYSIS

Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P190304-003
Expiration:	22 April 2023

Quality Control Method: Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

Specific Reaction: 2 – 4+ signal with the positive control and trace to 1+ signal with the negative control, no background. 0 – 50 infected cells per high power field.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 22 April 2019

Pattern Of Fluorescence: Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: Version # 2

FOR *IN VITRO* LABORATORY USE ONLY.

30 April 2019