

CERTIFICATE OF ANALYSIS

Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

| | |
|--------------|----------------|
| Catalog No.: | SLD-IFA-FIP1 |
| Lot: | P190812-001 |
| Expiration: | 13 August 2023 |

Quality Control Method: Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

Specific Reaction: 2-4+ signal with the positive control and trace with the negative control with no background. There are 0-50 infected cells per high-power field.

Other Comments: NA

QC Manager: 

David Hyatt

QC Release Date: 14 August 2019

Pattern Of Fluorescence: Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: Version # 2

FOR *IN VITRO* LABORATORY USE ONLY.

3 September 2019