

CERTIFICATE OF ANALYSIS

# Feline Infectious Peritonitis Virus Type 1 (FIP-1)


FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P190829-003
Expiration:	08 January 2024

**Quality Control Method:** Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-10ML).

**Specific Reaction:** 2-4+ signal with the positive control and negative with the negative control with trace background. There are 0-50 infected cells per high-power field.

**Other Comments:** NA

**QC Manager:** 

David Hyatt

**Pattern Of Fluorescence:** Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

13 January 2020