

CERTIFICATE OF ANALYSIS

Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P191216-002
Expiration:	23 January 2024

Quality Control Method: Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-10ML).

Specific Reaction: 2-4+ signal with the positive control and trace with the negative control with no background. There are 0-50 infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

17 February 2020