

CERTIFICATE OF ANALYSIS

# Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P200215-003(A)
Expiration:	23 March 2024

**Quality Control Method:** Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no. NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-10ML).

**Specific Reaction:** 2-4+ signal with the positive control and trace with the negative control with no background. There are zero to too numerous to count infected cells per high-power field.

**Other Comments:** NA

**QC Manager:**



David Hyatt

**Pattern Of Fluorescence:** Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

13 April 2020