

CERTIFICATE OF ANALYSIS

Feline Infectious Peritonitis Virus Type 1 (FIP-1)


FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P230223-001
Expiration:	04 April 2031

Quality Control Method: Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG affinity purified FITC conjugate (catalog no. CJ-F-FELG-AP-10ML).

Specific Reaction: 4+ signal with the positive control and negative with the negative control. No background. There are 0–40 infected cells per high-power field.

Other Comments: NA

QC Manager: 
David Hyatt

Pattern Of Fluorescence: Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

10 April 2023