

CERTIFICATE OF ANALYSIS

# Feline Herpesvirus Type 1 (FHV-1/FVR)

FA Substrate Slide

Catalog No.:	SLD-IFA-FVR
Lot:	P180205-001
Expiration:	16 March 2022

**Quality Control Method:** Indirect FA using FHV-1/FVR positive control (catalog no. PC-IFA-FVR), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

**Specific Reaction:** 2 – 4+ signal with the positive control and negative with the diluent control with no background. There are 0 - 20 infected cells and/or plaques per high-power field.

**Other Comments:** NA

**QC Manager:** Chris Cay



**QC Release Date:** 19 March 2018

**Pattern Of Fluorescence:** Individual rounded cells with cytoplasmic and nuclear fluorescence. Some small syncytia.

Technical Data Sheet Version: Version # 2

FOR *IN VITRO* LABORATORY USE ONLY.

02 April 2018