

CERTIFICATE OF ANALYSIS

Borrelia burgdorferi (Lyme Disease)

FA Substrate Slide

Catalog No.:	SLD-IFA-LD
Lot:	P160926-003
Expiration:	07 October 2020

Quality Control Method: Indirect FA using *Borrelia burgdorferi* positive control (catalog no. PC-IFA-LD), negative control (catalog no. NC-IFA-LD), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 3+ signal with the positive control, negative on the negative control with no background. There are greater than 25 organisms per high-power field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 07 October 2016

Pattern of Fluorescence: Spirochetes with uniform surface fluorescence.

Technical Data Sheet Version: Version # 2

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\8-, 10-, 12-well Slides\Lyme Disease\SLD-IFA-LD_P160926-003_201007.docx

12 October 2016