

CERTIFICATE OF ANALYSIS

Borrelia burgdorferi (Lyme Disease)

FA Substrate Slide

Catalog No.:	SLD-IFA-LD
Lot:	P221124-002
Expiration:	28 February 2031

Quality Control Method: Indirect FA using *Borrelia burgdorferi* positive control (catalog no. PC-IFA-LD), negative control (catalog no. NC-IFA-LD), and anti-canine IgG affinity purified FITC conjugate (catalog no. CJ-F-CANG-AP-10ML).

Specific Reaction: 2–4+ signal with the positive control and negative with the negative control. No background. Organisms per high-power field are too numerous to count.

Other Comments: NA

QC Manager:



Jennifer Boren

Pattern Of Fluorescence: Spirochetes with uniform surface fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

7 March 2023